

# MRI Room Accent Lighting Application Guide



Tempo has three low voltage linear LED product lines that are guaranteed to be compliant in MRI rooms


To Fully Comply With MRI Room Requirements:

- Lighting fixtures must be comprised of non-ferrous materials.
- All LED Drivers must be installed outside of the MRI room.
- Mount all Tempo luminaries using brass fasteners. (Part # C3A-MNTSCREW-BRS).
- Ferrous hardware is NOT permitted inside MRI rooms.

**CLiP 3 Dynamic**  
RGB, RGB-W, RGB-A and Tunable White.  
> [CLICK](#) for Product Specification Sheet



**C3R**  
White and Static color in rigid platform for straight runs.  
> [CLICK](#) for Product Specification Sheet



**C3S**  
White and Static color in serpentine platform for curved runs.  
> [CLICK](#) for Product Specification Sheet

