

# igScript™ Probe-Based qPCR 2x Master Mix



Catalog #	Package Size
4233	500 reactions
4235	1,000 reactions
4237	2,500 reactions

## Description

igScript™ Probe-Based qPCR 2x Master Mix contains igScript™ Taq DNA polymerase, MgCl<sub>2</sub>, dNTPs, stabilizers, enhancers and low ROX reference dye with standard buffer providing improved qPCR efficiency, wider dynamic range, superior sensitivity and specificity. igScript™ qPCR 2x Master Mix is a ready-to-use cocktail containing all components except primers, probe and template, for the amplification and detection of DNA in qPCR. This 2x master mix requires minimal handling during reaction setup and offer consistent and robust qPCR reactions. Taq DNA Polymerase is a thermostable DNA polymerase that possesses a 5'→3' polymerase (1, 2) and a 5'→3' exonuclease activity (3, 4). The amplification step features a high quality Taq DNA Polymerase which offers robust, reliable and better amplification.

## Product Includes

- igScript™ Probe-Based qPCR 2x Master Mix

**Storage Temperature:** -20 °C

## Applications

- Gene expression data validation.
- Multiplexing.
- Mutation detection.
- Pathogen and viral detection.
- Genetically modified organisms (GMO) characterization and Genetic profiling.

## Benefits

- Enhanced efficiency, specificity, and sensitivity
- Compatible with all real-time PCR instruments
- Superior gene expression results under various cycling conditions.

## Protocol

- Place kit components and DNA samples on ice.
- Mix and then centrifuge briefly to collect contents at the bottom of the tube.
- Prepare a master mix for each reaction and control plus 10% extra to allow for pipetting error, according to the following table:

PCR Reaction Set Up:	
Template DNA	x µl
Forward primer (5 µM)	1.0 µl
Reverse primer (5 µM)	1.0 µl
Probe (5 µM)	0.5 µl
igScript™ qPCR 2x Master Mix	10 µl
H <sub>2</sub> O up to	20.0 µl

- Mix the reaction mixture thoroughly.
- Program the thermal cycler according to the manufacturer's instructions.
- A typical PCR cycling program is outlined in the following table:

PCR Cycling Conditions			
Steps	Temperature	Time	Cycles
Initial denaturation	95°C	3 min	1
Denaturation	95°C	5 sec	40
Annealing/Extension*	55-60°C	30 sec	
Melting curve analysis	According to instrument guidelines		

- Place the PCR tubes in the thermal cycler and start the cycling program.
  - Analyze the data according to manufacturer protocol.
- \* For 3 step cycling protocols, anneal at optimal annealing temperature for 30 sec followed by the minimum time required for data acquisition at 72 °C according to instrument guidelines.

## References

- Chien, A., Edgar, D. B. and Trela, J. M. (1976). *J. Bact.* 127, 1550-1557.
- Lawyer, F. C. et al. (1993). *PCR Methods and Appl.* 2, 275- 287.
- Longley, M. J., Bennett, S. E. and Mosbaugh D. W. (1990). *Nucleic Acids Res.* 18, 7317-7322.
- Lyamichev, V., Brow, M. A. and Dahlberg, J. E. (1993). *Science.* 260, 778-783.

## Related Products

- igScript™ Probe Based One Step RT-qPCR Kit (Cat.#4243, 4245, 4247)
- igScript™ One Step RT-PCR Kit (Cat.# 4211)
- igScript™ One Step RT-qPCR Kit (Cat.# 4214)
- igScript™ First Strand cDNA Synthesis Kit(Cat.# 4312)
- igScript™ Reverse Transcriptase (Cat.# 3344)
- ig® SYBR Green qPCR 2x Master Mix (Cat.# 3354)

## Technical Support

Intact Genomics is committed to supporting the worldwide scientific research community by supplying the highest quality reagents. Each new lot of our products is tested to ensure they meet the quality standards and specifications designated for the product.

Please follow the instructions carefully and contact us if additional assistance is needed. We appreciate your business and your feedback regarding the performance of our products in your applications.

